



UNIVERSITÀ
di CAMERINO

Report

**Viral Filtration Efficiency (VFE) tests on sterile filters produced by the
Company Ethafilter s.r.l., Sovizzo (VI) Italy**

Laboratory Responsible

Prof. Alberto Cresci

Operator

Dr. Cinzia Cecchini

Scuola di Bioscienze e Biotecnologie

62032 Camerino (Italy)
Via Gentile III da Varano
Tel. +39 0737 403252
Fax +39 0737 403290

Direttore: claudio.gualerzi@unicam.it
Tel. +39 0737 403240

P.I. 00291660439
Cod. fisc. 81001910439
www.unicam.it



Results

The results obtained are shown in Table 1 expressed as mean values of three determinations $\pm$ standard deviations.

Table 1: Values of VFE % for the sterile filter element tested, expressed as mean values $\pm$ standard deviations.

Sample	Viral concentration tested (PFU) (mean $\pm$ sd)	Viral concentration downstream the filter (PFU) (mean $\pm$ sd)	Time (minutes)	VFE %
Sterile filter element	$1.00 \times 10^8 \pm 0.65 \times 10^7$	1.67 ± 0.58	2	99,999998
Sterile filter element	$1.11 \times 10^8 \pm 1.00 \times 10^7$	2.00 ± 1.00	15	99,999998

Responsible of research
Prof. Alberto Cresci